



**Request for Advice on Decontamination of Medical Devices**  
 Complete and return to  
[ggc.infectioncontroldecontamination@ggc.scot.nhs.uk](mailto:ggc.infectioncontroldecontamination@ggc.scot.nhs.uk)

Date	Requesting Person	Contact Details
14/08/2017	Alex Connell	0141 314-9575
Name of device	Lancer Storage Cabinet FD8 and FD8E	
Use (including type of patient and procedure)	Used to store endoscopes after decontamination and prior to patient use for 7 days.	
Reason for new device	N/A	
Manufacturer		Reusable / single use?
Lancer Uk Ltd		Reusable
Query	To increase storage time to 14 days to reduce the volume of emergency scopes being reprocessed due to scopes expiring.	
Details of manufacturers recommended decontamination		
Recommendation of Decontamination Group	It was agreed at the Decontamination Group and the Infection Prevention and Control SMT that the storage time could be increased to 14 days.	
Date	04/09/2017	