

Advice Request on Decontamination of Medical Devices

Complete and return to
ggc.infectioncontroldcontamination@ggc.scot.nhs.uk

Date	Requesting Person	Contact Details
22/08/2017	Graham Latta	g.latta@nhs.net
Name of device	Neonatal Transport Incubator	
Use (including type of patient and procedure)	Out of hospital transfer of neonates	
Reason for new device	Not a new device	
Manufacturer		Re-usable / Single Use?
Draeger		Re-usable
Query	Current recommendations for cleaning, following transfer of baby with CMV.	
Details of manufacturers recommended decontamination		
Recommendation of Decontamination Group	<p>Follow Standard Infection Control Precautions – Clean with detergent wipes unless any blood and body fluid spillages.</p> <p>http://www.nhsggc.org.uk/media/236969/decontamination-sop-v6-mar-2016.pdf</p>	

Date	04/09/2017
------	------------